
Niti-S SPAXUS™ Stent

for the drainage of pancreatic pseudocyst or a gallbladder through a transgastric or transduodenal approach



Lumen-apposing SPAXUS™ Stent

- Prevents migration and maintains lumen apposition
- Fully silicone coated preventing leakage and in-growth
- The flexible design helps accommodative apposition regardless of the wall thickness
- 8, 10, 16mm diameters enable to apply various indications

User friendly designed SPAXUS™ Delivery system

- Step1. When the outer X-ray marker overlaps with the inner X-ray marker
  → The distal flange is completely opened
- Step2. When the blue marker is visible under endoscopic view
  → Start opening proximal flange

TaeWoong MEDICAL

Ordering Information

Endoscopic Approach					
Code	Stent			Delivery	
	Body Diameter (mm)	Flare Diameter (mm)	Length (cm)	Profile (Fr)	Usable Length (cm)
SS0802FW	8	23	2	10	180
SS1002FW	10	25			
SS1602FW	16	31			

Distributed in the UK by:

Aquilant
A HC2I Company

csuk@hc21.group
01256 306506
healthcare21.eu

ⓒ 2022, Taewoong Medical Co., Ltd. All rights reserved. TW-MBEUS-S (Rev.00) Printed in Korea. May 04. 2022